



# K&S Associates, Inc.

1926 Elm Tree Drive  
Nashville, Tennessee 37210-3718  
800-522-2325 Fax 615-871-0856



## A2LA ACCREDITED INSTRUMENT CALIBRATION LABORATORY

FACILITY \_\_\_\_\_ PHONE \_\_\_\_\_

PERSON \_\_\_\_\_ FAX \_\_\_\_\_

INSTRUMENT \_\_\_\_\_

**A2LA Accredited calibration for Cesium-137 (0.5mR/h-2 R/h) 5 scales** \_\_\_\_\_  
Additional scales \_\_\_\_\_

High rate cost (2-1000 R/h), up to 3 scales/ranges \_\_\_\_\_

\*\*\*\*\*

### Additional Calibrations Available:

#### Exposure geometry:

- \_\_\_\_\_ Ir-192 photons (not accredited)
- \_\_\_\_\_ I-125 photons (not accredited)
- \_\_\_\_\_ Pd-103 (not accredited)
- \_\_\_\_\_ 70kVp x-rays (AAPM Accredited)
- \_\_\_\_\_ 90kVp x-rays (AAPM Accredited)
- \_\_\_\_\_ 110kVp x-rays (AAPM Accredited)

\*\*\*\*\*

#### Contamination geometry: (not accredited)

##### Beta emitters:

- \_\_\_\_\_ C-14 (156 KeV)
- \_\_\_\_\_ Pm-147 (225 KeV)
- \_\_\_\_\_ Tc-99 (294 KeV)
- \_\_\_\_\_ Cl-36 (1142 KeV)
- \_\_\_\_\_ Sr90/Y90 (546, 2282 KeV)

##### Gamma emitters:

- \_\_\_\_\_ I-129 photon source (I-125 mimic)

##### Alpha emitters:

- \_\_\_\_\_ Pu-239 alpha

\*\*\*\*\*

\*In addition to basic calibration on Cesium -137

Note: Electronic linearity is performed on all instruments reading in CPM units

### Other Calibration Services:

- Pocket dosimeters (0-200 mR) \_\_\_\_\_
- Electronic integrating personal beepers \_\_\_\_\_
- Radiation monitor/beepers \_\_\_\_\_
- Unit unfit/rejected for calibration \_\_\_\_\_

**Calibration intervals: Your re-calibration reminder will be set on a 12 month interval unless you specify otherwise. IF you wish a re-calibration date recorded on your report please indicate the interval you require in the space provided.** \_\_\_\_\_

*American Express, MasterCard, Visa are accepted* email: ksinfo@kslab.com web: <http://www.kslab.com>